IAES Journal of Electrical and Electronics Engineering
Vol. 2, No. 1, June 2015, pp. 6~10
ISSN: 2338-7432, DOI: 10.11591/jeee.v2i1.pp6-10

Wearable sensors for health monitoring and disease prevention: a review of technologies and applications

M. Nasif Averus[1], Lina Handayani[2]
[1]Faculty of Industrial Technology, Universitas Ahmad Dahlan, Yogyakarta, Indonesia
[2]Faculty of Public Health, Universitas Ahmad Dahlan, Yogyakarta, Indonesia

Article Info

Article history:

Received May 16, 2014
Revised Jan 16, 2015
Accepted May 8, 2015

Keywords:

Disease prevention
Health monitoring
Physiological monitoring
Preventive healthcare
Remote patient monitoring
Wearable sensors
Wireless body area networks

ABSTRACT

Wearable sensors have emerged as important technologies for continuous health monitoring, early detection of physiological abnormalities, and preventive healthcare. These devices integrate miniaturized sensors, electronic components, wireless communication, and data processing capabilities to collect physiological and behavioral information during daily activities. Common measurements include heart rate, electrocardiogram (ECG), blood pressure, body temperature, oxygen saturation, physical activity, and respiratory parameters. This review examines the fundamental technologies, system architectures, data processing methods, and healthcare applications of wearable sensors, emphasizing developments documented up to 2013. Particular attention is given to cardiovascular monitoring, chronic disease management, physical activity assessment, elderly care, and remote patient monitoring. The review also discusses challenges involving measurement accuracy, sensor reliability, energy consumption, data security, and clinical validation. Overall, wearable sensors provide opportunities to strengthen preventive healthcare and expand access to continuous health monitoring, although their effectiveness depends on reliable measurements, appropriate clinical interpretation, and integration with healthcare services.

This is an open access article under the CC BY-SA license.

1. INTRODUCTION

The increasing prevalence of chronic diseases and the growing demand for accessible healthcare have encouraged the development of technologies capable of monitoring human health beyond conventional clinical environments. Traditional healthcare assessments generally depend on periodic examinations, which may not capture physiological changes occurring during everyday activities. Wearable sensors provide an alternative by enabling continuous or repeated measurements of physiological and behavioral parameters in relatively unobtrusive ways.

Wearable health monitoring systems combine sensing devices, embedded electronics, communication technologies, and analytical algorithms to collect and interpret health-related information. These systems can monitor cardiovascular activity, body movement, respiratory conditions, temperature, and other physiological characteristics. Pantelopoulos and Bourbakis [1] demonstrated the potential of wearable sensor-based systems to support proactive personal health management and continuous monitoring.

The application of wearable sensors extends beyond individual health assessment to preventive healthcare and population health management. Continuous monitoring may facilitate earlier identification of abnormal physiological patterns, improve chronic disease management, and support remote healthcare delivery. This review examines the technologies, system architectures, applications, and limitations of wearable sensors based exclusively on scientific literature published no later than 2013.

2. WEARABLE SENSOR TECHNOLOGIES

Wearable health monitoring systems incorporate different sensing technologies depending on the physiological parameters and health conditions being investigated. These technologies transform physical,

electrical, optical, or biochemical phenomena into measurable signals that can be processed for health assessment. Their selection depends on measurement accuracy, sensor placement, user comfort, power consumption, and intended clinical applications.

2.1. Physiological sensors

Physiological sensors measure parameters associated with the functioning of major human body systems. Electrocardiogram (ECG) sensors record cardiac electrical activity, while photoplethysmography sensors measure changes in peripheral blood volume to estimate pulse-related characteristics. Temperature sensors, respiratory sensors, and blood pressure measurement devices provide additional information about physiological conditions.

These sensors can be integrated into wearable devices, clothing, adhesive patches, and portable monitoring equipment. Their combination allows multiple physiological parameters to be measured within a single monitoring platform. Baig *et al.* [2] reviewed wearable and wireless ECG systems and highlighted their potential for continuous cardiovascular monitoring, particularly among older adults.

2.2. Motion and activity sensors

Motion sensors provide information about physical activity, body movement, posture, gait, and mobility. Accelerometers measure changes in linear acceleration, while gyroscopes provide information about angular velocity and rotational movement. Combining these sensors can improve the characterization of physical activities and movement patterns.

Motion monitoring is particularly relevant for elderly care, rehabilitation, fall detection, and assessment of physical activity. Patel *et al.* [3] discussed the application of wearable sensors in rehabilitation and home monitoring, including the assessment of functional mobility and treatment outcomes. These technologies can also support preventive health programs by providing information about daily activity patterns and sedentary behavior.

2.3. Wireless body area networks

Wireless body area networks (WBANs) connect wearable sensors and processing devices through short-range wireless communication technologies. A typical architecture consists of body-mounted sensors, a local processing or gateway device, and a remote monitoring platform. This arrangement enables physiological measurements to be collected and transmitted for further analysis without requiring permanent physical connections to clinical equipment.

Bluetooth, ZigBee, and other wireless communication technologies have been investigated for wearable healthcare applications. Alemdar and Ersoy [4] reviewed wireless sensor networks for healthcare and discussed their potential for continuous monitoring and pervasive health services. Custodio *et al.* [5] further examined wearable health monitoring architectures and communication technologies, highlighting the importance of reliable connectivity and system integration.

3. SYSTEM ARCHITECTURE AND DATA PROCESSING

The effectiveness of wearable health monitoring depends on the integration of sensing, communication, processing, and information management components. A complete system must acquire reliable measurements, transmit information efficiently, identify relevant physiological patterns, and communicate useful results to users or healthcare professionals. These components must operate together to provide meaningful monitoring rather than simply collecting large volumes of sensor data.

3.1. Data acquisition and signal processing

Data acquisition is the initial stage of wearable health monitoring and involves collecting physiological and motion measurements through appropriate sensors. The acquired signals may contain noise, motion artifacts, missing observations, and variations associated with individual characteristics. Signal preprocessing, filtering, normalization, and segmentation are therefore necessary to improve data quality before further analysis.

Different physiological parameters require different processing approaches. ECG signals may require waveform detection and heart rhythm analysis, while accelerometer measurements can be used to identify movement patterns and physical activities. Proper processing is essential because measurement errors can generate false alarms or conceal potentially important physiological abnormalities.

3.2. Feature extraction and intelligent analysis

Feature extraction transforms raw sensor measurements into representative characteristics that can be used for health assessment. Examples include heart rate, heart rate variability, movement intensity, activity duration, respiratory patterns, and changes in physiological measurements over time. These features provide a more interpretable representation of health conditions than raw sensor signals alone.

Data mining and machine learning techniques can subsequently be applied to identify patterns, detect abnormalities, and support decision-making. Banaee *et al.* [6] reviewed anomaly detection, prediction, and classification methods for wearable health monitoring systems. Their review demonstrated the importance of analytical methods in transforming continuous sensor measurements into potentially useful healthcare information.

3.3. Remote monitoring and information communication

Remote monitoring systems transmit wearable sensor measurements to healthcare professionals or monitoring centers through wireless communication networks. These systems can support continuous observation of individuals without requiring frequent visits to healthcare facilities. Suh *et al.* [7] demonstrated this approach through a remote patient monitoring system designed for individuals with congestive heart failure.

Information communication is also important for generating alerts when physiological measurements exceed predefined thresholds or exhibit unusual patterns. However, alert mechanisms must be carefully designed to minimize false alarms and avoid unnecessary clinical interventions. Effective remote monitoring therefore requires appropriate sensor performance, communication reliability, data interpretation, and clinical response procedures.

4. APPLICATIONS IN HEALTH MONITORING AND DISEASE PREVENTION

Wearable sensors have been investigated across several healthcare domains because continuous physiological and behavioral measurements can provide additional information about individual health conditions. Their applications include cardiovascular monitoring, chronic disease management, physical activity assessment, elderly care, and remote patient monitoring. These applications demonstrate the potential of wearable technologies to complement conventional healthcare services.

4.1. Cardiovascular health monitoring

Cardiovascular monitoring is one of the most established applications of wearable sensor technologies. ECG and PPG sensors can provide information about cardiac electrical activity, pulse rate, and changes in cardiovascular function. Continuous measurements may help identify irregular heart rhythms and other physiological changes that require further clinical assessment.

Wearable and wireless ECG systems are particularly relevant for individuals with cardiovascular risk factors and chronic cardiac conditions. Baig *et al.* [2] examined the technological developments and applications of wearable ECG systems for older adults. Such systems may improve monitoring outside hospitals and provide additional information for evaluating cardiovascular health over time.

4.2. Chronic disease management

Wearable monitoring technologies can support the management of chronic diseases by providing repeated physiological measurements during daily activities. Individuals with cardiovascular disorders, respiratory conditions, and metabolic diseases may benefit from monitoring selected health parameters outside clinical environments. Continuous information can help healthcare professionals evaluate changes in patient conditions and assess responses to treatment.

Suh *et al.* [7] presented a remote monitoring system for congestive heart failure that integrated physiological measurements and communication technologies. This application illustrates how wearable and remote monitoring approaches can support chronic disease management. Nevertheless, monitoring technologies should complement established medical assessment and treatment rather than independently determine clinical decisions.

4.3. Physical activity and preventive healthcare

Physical activity monitoring is an important application of wearable sensors because physical inactivity and sedentary behavior are associated with several chronic health risks. Accelerometers and other motion sensors can measure activity intensity, duration, and movement patterns during everyday activities. These measurements may provide feedback that encourages healthier behavioral patterns and supports preventive health programs.

Wearable activity monitoring can also support the assessment of individual responses to exercise interventions and lifestyle modifications. Patel *et al.* [3] discussed the role of wearable sensors in health and wellness applications, rehabilitation, and monitoring of treatment effectiveness. However, sensor measurements alone cannot establish whether an individual has achieved meaningful improvements in long-term health outcomes.

4.4. Elderly care and fall detection

Wearable sensors offer opportunities to improve health monitoring and independent living among older adults. Accelerometers, gyroscopes, and physiological sensors can be used to assess mobility, physical activity, and potentially dangerous movement events. These capabilities may support fall detection, mobility assessment, and timely communication with caregivers.

Remote monitoring can also provide information about changes in daily activity patterns that may indicate declining functional capacity. Patel *et al.* [3] emphasized the potential of integrating wearable and environmental sensors for monitoring older adults and individuals with chronic conditions in home and community settings. Nevertheless, fall detection systems require careful validation because normal activities can sometimes produce movement patterns similar to falls.

4.5. Remote and community healthcare

Wearable monitoring systems can extend healthcare services to individuals who live far from hospitals or have difficulties accessing conventional healthcare facilities. Portable sensors and wireless communication technologies enable health-related measurements to be collected in homes and community environments. This capability is particularly relevant for remote patient monitoring and community-based chronic disease management.

Khusainov *et al.* [8] reviewed real-time human ambulation, activity, and physiological monitoring technologies, identifying applications and challenges associated with continuous monitoring. These developments demonstrate the potential of wearable technologies to support more accessible healthcare services. However, effective community implementation requires appropriate communication infrastructure, affordable devices, trained healthcare personnel, and reliable clinical referral mechanisms.

5. CHALLENGES AND LIMITATIONS

Despite their potential benefits, wearable health monitoring systems face several technical, clinical, and organizational challenges. Measurement accuracy can be affected by sensor positioning, movement artifacts, environmental interference, skin contact, and individual physiological differences. These factors may reduce the reliability of collected information, particularly during prolonged monitoring and unrestricted daily activities.

Energy consumption and communication reliability represent additional engineering challenges. Continuous sensing and wireless transmission can increase battery requirements, while unstable network connections may result in missing or delayed information. System designers must therefore balance measurement frequency, processing requirements, communication performance, and device operating time.
Data privacy, security, and user acceptance are equally important considerations. Wearable systems may collect sensitive physiological and behavioral information that requires appropriate protection and access management. Furthermore, devices must be comfortable, affordable, and sufficiently simple to encourage long-term use.

A particularly important limitation concerns the clinical interpretation and generalizability of wearable sensor measurements. Algorithms developed using limited datasets may not perform consistently across different populations, age groups, or real-world conditions. Therefore, independent clinical validation and evaluation of long-term health outcomes are necessary before wearable technologies can be widely adopted as reliable disease prevention tools.

6. FUTURE DIRECTIONS

Future development of wearable health monitoring systems should emphasize the integration of multiple physiological and motion sensors into reliable and energy-efficient platforms. Advances in embedded processing, wireless communication, and intelligent data analysis may improve the ability to identify meaningful physiological changes in real time. Multisensor integration can also provide more comprehensive information about individual health conditions than isolated measurements.

Another important direction is the integration of wearable monitoring with electronic health records and public health information systems. Such integration could support longitudinal health assessment, remote patient management, and more coordinated preventive healthcare services. However, these developments

require standardized data exchange, secure information management, and clear procedures for interpreting and responding to monitoring results.

Research should also prioritize clinical validation, affordability, user acceptance, and equitable access. Wearable technologies will have greater public health value when they can be implemented reliably in diverse healthcare environments rather than being restricted to technologically advanced settings. Consequently, future innovation should address both engineering performance and practical healthcare requirements.

7. CONCLUSION

Wearable sensors provide important technological capabilities for continuous health monitoring, early identification of physiological abnormalities, and preventive healthcare. Physiological sensors, motion sensors, WBANs, and intelligent data processing enable the development of monitoring systems for cardiovascular health, chronic disease management, physical activity, elderly care, and remote healthcare delivery. These applications demonstrate the potential of wearable technologies to extend health monitoring beyond conventional clinical environments. Nevertheless, their effectiveness depends on measurement accuracy, reliable communication, appropriate data interpretation, clinical validation, and user acceptance. Technical limitations, energy consumption, privacy concerns, affordability, and unequal access must also be addressed. Overall, wearable sensors should be viewed as complementary technologies that can strengthen preventive healthcare, support continuous health assessment, and improve the accessibility of population-oriented healthcare services.

REFERENCES

[1] A. Pantelopoulos and N. G. Bourbakis, "A survey on wearable sensor-based systems for health monitoring and prognosis," *IEEE Transactions on Systems, Man and Cybernetics Part C: Applications and Reviews*, vol. 40, no. 1, pp. 1–12, 2010, doi: 10.1109/TSMCC.2009.2032660.

[2] M. M. Baig, H. Gholamhosseini, and M. J. Connolly, "A comprehensive survey of wearable and wireless ECG monitoring systems for older adults," *Medical and Biological Engineering and Computing*, vol. 51, no. 5, pp. 485–495, 2013, doi: 10.1007/s11517-012-1021-6.

[3] S. Patel, H. Park, P. Bonato, L. Chan, and M. Rodgers, "A review of wearable sensors and systems with application in rehabilitation," *Journal of NeuroEngineering and Rehabilitation*, vol. 9, no. 1, 2012, doi: 10.1186/1743-0003-9-21.

[4] H. Alemdar and C. Ersoy, "Wireless sensor networks for healthcare: a survey," *Computer Networks*, vol. 54, no. 15, pp. 2688–2710, 2010, doi: 10.1016/j.comnet.2010.05.003.

[5] V. Custodio, F. J. Herrera, G. López, and J. I. Moreno, "A review on architectures and communications technologies for wearable health-monitoring systems," *Sensors (Switzerland)*, vol. 12, no. 10, pp. 13907–13946, 2012, doi: 10.3390/s121013907.

[6] H. Banaee, M. U. Ahmed, and A. Loutfi, "Data mining for wearable sensors in health monitoring systems: A review of recent trends and challenges," *Sensors (Switzerland)*, vol. 13, no. 12, pp. 17472–17500, 2013, doi: 10.3390/s131217472.

[7] M. K. Suh *et al.*, "A remote patient monitoring system for congestive heart failure," *Journal of Medical Systems*, vol. 35, no. 5, pp. 1165–1179, 2011, doi: 10.1007/s10916-011-9733-y.

[8] R. Khusainov, D. Azzi, I. E. Achumba, and S. D. Bersch, "Real-time human ambulation, activity, and physiological monitoring: Taxonomy of issues, techniques, applications, challenges and limitations," *Sensors (Switzerland)*, vol. 13, no. 10, pp. 12852–12902, 2013, doi: 10.3390/s131012852.